

**Thickness of lead Required for a Primary
Beam Barrier Located 5 cm from the Focal Spot**

Anode Current (ma)	Thickness of lead (mm)		
	50kVp	70kVp	100kVp
20	1.5	5.6	7.7
40	1.6	5.8	7.9
80	1.6	5.9	
160	1.7		