

Figure: 25 TAC §289.232(j)(5)(J)(ii)

Type of Machine	Frequency
CBCT	Annually not to exceed 14 months from the prior EPE
All other Dental X-ray	Four years from the date of prior EPE