

Figure: 25 TAC §289.230(v)(5)(D)

TABLE I

X-ray Tube Voltage (kilovolt peak) and Minimum HVL

Designed Operating Range (kV)	Measured Operating Voltage (kV)	Minimum HVL (mm of aluminum)
Below 50	20	0.20
	25	0.25
	30	0.30