

Figure: 25 TAC §289.230(ee)(3)

Specific Subsection	Name of Record	Time Interval for Record Keeping
(r)(1)(A)	Interpreting Physician Qualifications	Until termination of certification or 2 years after physician leaves facility
(r)(1)(C)	Interpreting Physician Continuing Education and Experience	6 years
(r)(1)(E)	Mandatory training for Interpreting Physician, if applicable	6 years
(r)(2)(A)	Medical Radiologic Technologist Qualifications	Until termination of certification or 2 years after technologist leaves facility
(r)(2)(C)	Medical Radiologic Technologist Continuing Education and Experience	6 years
(r)(2)(E)	Mandatory training for Medical Radiologic Technologist, if applicable	6 years
(r)(3)(A)	Medical Physicist Qualifications	Until termination of certification or 2 years after physicist is no longer associated with the facility
(r)(3)(C)	Medical Physicist Continuing Education and Experience	6 years
(s)(14)	FDA Variances	Until termination of certification or equipment is replaced
(u)(2)	Quality Assurance (QA) Records	Until the next annual inspection has been completed and the agency has determined that the facility is in compliance with the QA requirements or until the test has been performed two additional times at the required frequency, whichever is longer.
(v)(10)	Physicist Mammography Survey	7 years
(v)(11)	Physicist Mammography Equipment Evaluation	2 years
(w)(2)	Medical Outcomes Audit	2 years
§289.234(o)	Complaints	3 years
(dd)(1)	Operating & Safety Procedures	Until termination of certification
(dd)(3); (gg)(15)	Records of Receipt, Transfer, and Disposal	Until termination of certification
(dd)(6)	Protective Devices Annual Check	3 years
(dd)(8)	Records on Calibration, Maintenance and Modifications Performed on Mammography Machines	2 years
(ee)(2)(I)	Current §§289.203, 289.204, 289.205, 289.226, 289.227, 289.230, 289.231, 289.234	Until termination of certification
(ee)(2)(J)	Current Certification of Mammography Systems	Until termination of certification

(ee)(2)(N)	Current Accreditation of Mammography Systems	Until termination of certification
(ff)(5)	Certification of Inspection	Until termination of certification
(ff)(6)	Notice of Failure	Until termination of certification
(ff)(10)	Patient Notification	Until termination of certification
(gg)(16)	Records of Calibration, Maintenance, and Modifications Performed on Interventional Breast Radiography Machines	Until termination of certification