

Figure: 25 TAC §289.231(II)(2)

Machine Type/Type of Use	Years Between Inspections
CT	2 years
Fluoroscopy	2 years
Accelerators, Simulators, and Other Therapeutic Radiation Machines	2 years
Radiographic Only	3 years
Podiatric Radiographic Only	4 years
Minimal Threat	5 years
Industrial Radiography	1 year
Other Industrial Services	5 years
Laser (Human Use/Research/Academic)	As deemed necessary by the agency
Other Laser	As deemed necessary by the agency
Mammography	1 year

NOTE: The inspection intervals specified above were based upon the average number of health-related violations per inspection by category, as determined from compliance history data. These intervals will be reviewed at least every 2 years, and appropriate adjustments will be made.